

# NASA TECH BRIEF



NASA Tech Briefs are issued to summarize specific innovations derived from the U.S. space program, to encourage their commercial application. Copies are available to the public at 15 cents each from the Clearinghouse for Federal Scientific and Technical Information, Springfield, Virginia 22151.

## Direct Reading of Electrocardiograms and Respiration Rates




Fig. 1. Scaled card




Fig. 2. Card indicating 19 respirations per minute

### The problem:

To develop a fast, accurate method of reading heart and respiration rates. Formerly, the quick method of reading recorder traces of electrocardiograms and respiration rates consisted in analysis of a 10-second period and multiplication by six for a count; errors could result.

### The solution:

A new technique enables quick direct reading of heart and respiration rates. The appropriate index of a calibrated card of transparent plastic, about 6

inches long (Fig. 1), is aligned with a repetitive point on the electrocardiogram or respiration complex. For recorder speeds of 10 and 25 mm/sec the number of complexes are counted as indicated on the card, and the heart or respiration rate is read directly from the appropriate scale (Fig. 2); for other recorder speeds, the number of complexes to be counted is determined by simple numerical conversion (speed 5 mm/sec = 10 ECG, 2 ZPN). The readings yielded are accurate as well as immediate; they save much time in data analysis.

(continued overleaf)

Also incorporated on the card is a second indicator for aiding measurement of time intervals on the electrocardiogram complex at 25 mm/sec. Doctors, hospitals, and health authorities may be interested.

**Note:**

No further documentation is available. Inquiries may be directed:

Technology Utilization Officer  
Kennedy Space Center  
Kennedy Space Center, Florida 32899  
Reference: B69-10188

**Patent status:**

Inquiries about rights for commercial use of this invention may be made to NASA, Code GP, Washington, D.C. 20546.

Source: J. P. Wise  
Kennedy Space Center  
(KSC-10233)